

# OLYMPIC PATIENT MOVER

**Product Code:**

HLA046

**Description:**

The Olympic Mover is an attractive universal-fit resident transport solution, providing practicality alongside a home-from-home aesthetic. The mover is designed for residents that cannot move without assistance but have some ability to independently act or contribute to pulling themselves upright. The robust, ergonomic design improves resident safety and gives peace of mind for carers, while the rotating well-padded seats and deep kneepad ensure optimum comfort. User-friendly total-lock castors and leveraged leg spread allow easy use even on a thick carpet, making the Olympic Mover safe and efficient for carers as well as residents. This product is suitable for on-site transport and transfer of residents, allowing carers to provide a calm and dignified mobility solution.

**Safe Working Load:**

180kg (28 Stone)

**External Width:**

1060mm (Legs Open)

**Footprint (L x W)**

1500x650mm


**Turning Radius:**

1125mm

**Features:**

- Provides easy transportation to the bathroom and toilet
- Offers a suitable alternative to a wheelchair
- Allows for better eye contact between the user and carer
- Padded knee support for extra comfort
- Manual leg open & closing

## Optional Extras



### Mackworth Back Support Belt

- Available in S, M, L, XL
- Product Codes:  
1201TAB510C: Mackworth Back Support Belt S  
1201TAB520C: Mackworth Back Support Belt M  
1201TAB530C: Mackworth Back Support Belt L  
1201TAB540C: Mackworth Back Support Belt XL

### Warranty:

Subject to the terms of Harvest Healthcare Guarantee and Warranties the Olympic Mover is guaranteed for one year from date of purchase.

### Supplier Contact Details:

t: +44(0)1709 377172 e: [sales@harvesthealthcare.co.uk](mailto:sales@harvesthealthcare.co.uk); w: [www.harvesthealthcare.co.uk](http://www.harvesthealthcare.co.uk)  
Sheaf house, Bradmarsh Way, Bradmarsh Business Park, Rotherham UK, S60 1BW

#### Important Safety Notice

Harvest Healthcare Ltd supplies equipment and appliances to cover a range of medical conditions. Specific choice of equipment is the sole responsibility of the clinical practitioner or carer overseeing individual patient care. Risk assessments specific to user must be carried out prior to the use of all products. Products and data continually undergo development and may therefore deviate from the details given.

May 2021