



User Manual

EASYFIT UNIVERSAL SLING YEW RANGE

INSTRUCTIONS FOR USE

Symbols in use



Read all warnings in the Instructions for use



Manufacturers details



Prior to use: -

- Ensure you are aware of the hoist manufacturer's fitting instructions
- Inspect the sling for wear and damage, with particular attention being paid to the body of the sling, webbings and stitch lines.
- If the sling is fitted with a head support, ensure the 2 head stays are fitted in the head pockets.
- If following inspection, there is any doubt, (points 1-3) the sling should be placed aside and not used, then reported to the matron.

User Instruction

How to fit from a seated position

1. Place the sling behind the patient ensuring that all handles and labels are on the outside and visible. Feed the sling down the back of the patient. This should be at the base of the spine, and positioned centrally to the patients back. The sling should now lie square across the patient's shoulders. If you find this is either higher or lower than the shoulder line, the back of the sling may not have been positioned deep enough, or the sling size may need to be considered.
2. Raise each leg individually and gradually feed the leg straps under each leg, ensuring no fabric twists are evident. Taking the inner strap from under one leg feed the webbing through the lower loop on the other inner leg strap. By carrying out this arrangement, you are ensuring comfort for the patient during the lifting process.
3. With the sling now in place, attach the desired loops to the hooks on the spreader bar to replicate the sitting position that the patient is in whilst seated, the straps should be positioned accordingly, subsequently the tension on all six straps should take at the same time
4. Make sure that the loops selected are the same on each side of the sling and ensure that they are still securely attached before lifting the patient. The person can now be lifted as required. To remove the sling, carry out the fitting procedure in reverse. Important – always peel the sling away from the person – this avoids skin friction.

How to fit from a lying position

1. Using good manual handling techniques roll the person towards you so that they are positioned on their side. To introduce the sling, roll half of the sling up under itself so that when the person is rolled onto their opposite side, the sling can be unrolled until it is flat. Use the centre pocket to ensure the base of the back section is in line with base of the person's spine. Ensure that handle loops are to the outside.
2. Raise each leg individually and gradually feed the leg straps under each leg, ensuring no fabric twists are evident. Taking the inner strap from under one leg feed the webbing through the lower loop on the other inner leg strap. By carrying out this arrangement, you are ensuring comfort for the patient during the lifting process.
3. The six straps should now be attached to the spreader bar hooks. Raise the carry bar just enough to tension the straps and Make sure that the loops selected are the same on each side of the sling and ensure that they are still securely attached before lifting the patient. The person can now be lifted as required. To remove the sling, carry out the fitting procedure in reverse. Important – always peel the sling away from the person – this avoids skin friction.



Please note: To avoid damage in laundering, please ensure any Velcro products are closed prior to washing.

The safe working load and serial numbers are embroidered onto the slings centre tab – **DO NOT** use this tab as a handle or for any lifting assistance.

Washing Instructions:

- The sling can be machine or hand washed in a mild soap solution
- Temperatures up to 85°C with a final rinse.
- If tumble drying is necessary – use a low heat setting.
- **DO NOT** use any corrosive cleaning agents on the sling to clean

Contra-indications

There are no contra indications associated with the use of the ranges of Slings manufactured. The sling is transient in its use and only in contact with uncovered skin for a very short period (5 to 10 minutes' maximum), the sling is not an active device and is non-invasive.

Cautions

There are no user medical cautions but as will all devices, the IFU should be read at all times and the limitations of the device understood regarding the usability of the slings.

Intended use

This is derived through the information on the previous pages.



For the safety of the patient and the carer, before using the sling, a full risk assessment must be conducted to ensure that the correct sling choice, method of positioning in the sling and procedure for transfer has been determined for the patient.
Check sling before each use, bleached, torn, cut, frayed or broken slings are unsafe and could result in serious injury or death to the patient.

**DESTROY AND DISCARD WORN SLINGS / DO NOT ALTER SLINGS
NEVER LEAVE A PATIENT UNATTENDED**

PLEASE ENSURE THE SLING, IF LEFT CONNECTED TO THE HOIST SYSTEM IN USE, IS SECURED TO THE SYSTEM WHEREBY ANY MINOR CANNOT INADVERTENTLY INJURE THEMSELVES.

PLEASE STORE THE SLING IN A RESPONSIBLE MANNER TO ENSURE IT HAS NO MEANS OF BEING DAMAGED WHILST NOT IN USE OR HAS THE ABILITY TO BECOME A MEANS OF AN INTRODUCTION OF RISK FOR UNINTENDED USERS SUCH AS MINORS PLAYING ON THE SLING AND THE SYSTEM IT IS ATTACHED TO.

Operating environment:

The slings are textile based and contain no electrical / electronic installed systems. The slings are used, and can only be used, within the same working environment as the mobile and ceiling hoists to which they are an accessory for use with.

Standards Applied

The following standards are applied to the product:

- BS EN ISO 10535 – Hoists for the transfer of disabled persons – Requirements and Test Methods
- LOLER - Lifting Operations and Lifting Equipment Regulations” (LOLER 5 Dec 1998),



Prism Medical UK

Unit 1, Tir Llwyd Industrial Estate, St Asaph Avenue, Kinmel Bay, Conwy, LL18 5JZ

Tel: +44(0)1924 840100

email: info@prismmedical.co.uk

